Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Linezolid LNZ 0.016-256	042826038	2029.04.27

Ref.

92135 – 921350 – 921351

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (µg/mL)	Results MIC(µg/mL)
Staphylococcus aureus ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4 [a,b]	1
Enterococcus faecalis ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4 [a,b]	2
Staphylococcus aureus ATCC® 25923	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4 [a]	2
Streptococcus pneumoniae ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO_2 20-24 h	0.25–2 [a,b]	1

a) CLSI M100-Ed35

b) EUCAST Quality Control v.16.0

Batch Release

Approved

Date 05.05.2026 **Signature** Quality Control

(This document has been established electronically and is valid without signature)

The results reported were obtained at the time of release.