Liofilchem®

Certificate of Analysis

Product	Batch	Expiration date
Ceftolozane-tazobactam C/T 0.016/4-256/4	101526031	2028.10.14

Ref.

92146 – 921460 – 921461

Antimicrobial Susceptibility Testing

Tested according to current CLSI and/or EUCAST methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (µg/mL)	Results MIC (µg/mL)
Escherichia coli ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.12–0.5 [a,b]	0.25
Pseudomonas aeruginosa ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.25–1 [a,b]	0.5
Escherichia coli ATCC® 35218	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.06–0.25 [a,b]	0.125
Klebsiella pneumoniae ATCC® 700603	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.5–2 [a,b]	1

a) CLSI M100-Ed35

b) EUCAST Quality Control v.16.0

Batch Release

Approved

Date 22.10.2026 **Signature** Quality Control

(This document has been established electronically and is valid without signature)

The results reported were obtained at the time of release.