www.pro-lab.com

CERTIFICATE OF ANALYSIS

Product Name:	**Prolex™ Staph Xtra Latex Kit**	**Lot Number:**	**C2399**
Product Code:	**PL.1080**	**Expiration Date:**	**2027-08-31**

Product Code	Quantity	Components
PL.1083	2 x 2.5 ml	Staph Xtra Latex Reagent
PL.1085	1 x 5.0 ml	Staph Xtra Negative Control Latex Reagent
	200	Mixing Sticks
	20 cards	Test Cards

Additional Information

Storage Conditions:	2-8ºC
Instructions for Use:	Refer to pro-lab.com/resources for current version

Quality Control Testing

Criteria	Result
Strong agglutination of Staph Xtra Latex Reagent when tested against *Staphylococcus aureus* and MRSA* including capsular types 5 and 8 within 20 seconds.	PASS
No agglutination of Staph Xtra Negative Control when tested against test isolates at 20 seconds.	PASS

* MRSA - (Methicillin-resistant *Staphylococcus aureus*)

Release

2026-02-13	
Date	Quality Control

Pro-Lab Diagnostics U.K.
3 Bassendale Road, Bromborough, Wirral,
Merseyside CH62 3QL
United Kingdom
Tel.: 0151 353 1613
www.pro-lab.co.uk
E-mail: uksupport@pro-lab.co.uk

Pro-Lab Diagnostics Canada
20 Mural Street, Unit #4
Richmond Hill, Ontario, L4B 1K3
Canada
Tel.: 905-731-0300
www.pro-lab.com
E-mail: support@pro-lab.com

Pro-Lab Diagnostics U.S.A.
1301 Blue Ridge Drive, Suite #101
Georgetown, Texas 78626-3269
United States
Tel.: 512-832-9145
www.pro-lab-direct.com
E-mail: support@pro-lab.us