CERTIFICATE OF ANALYSIS
SI 390.900 – URO-QUICK SCREENING KIT

Code MIC_COA_SI390900#SCR260319

PRODUCT NAME	URO-QUICK SCREENING KIT
REF NUMBER	SI 390.900
LOT	SCR260319
EXPIRATION DATE (yyyy-mm-dd)	2027-11-29

Tests	Specifications	Results
Visual inspection		
Appearance of the liquid culture medium *Appearance of the vial*	*Limpid, yellow, transparent liquid* *Green ring cap, pierceable septum, white stripe, presence of the stirrer, absence of defects*	OK
Microbiological activity		
Escherichia coli ATCC 8739	*Growth detection with inoculum ≤ 100 CFU*[1]	OK
Microbiological state		
72 hours of incubation at 37°C	*Absence of turbidity of the liquid culture media*[2]	OK

[1]Inoculum concentration guaranteed using certified standards.
[2]Visual inspection of the liquid culture medium after incubation

The sampling method and quality controls have been performed in accordance with the procedures in force.

Analysis outcome:	☑ Compliant ☐ Not compliant

Lot approved on	2026-10-05	**By:** (Quality Assurance)	Elena Lorenzo

©Copyright [2025]. This document is exclusive property of ALIFAX S.r.l, confidential and not freely disclosable. All rights reserved.

ALIFAX S.r.l.
Registered and head office
via Petrarca, 2/1 - 35020 Polverara PD Italy
Tel. (+39) 049 0992000 Fax (+39) 049 5855434
info@alifax.com - info.alifax@pec.alifax.com

R&D, Design and Manufacturing Dept
via Merano, 30,33045 Nimis UD Italy
Tel. (+39) 0432 197900 Fax (+39) 0432 547378
info.udine@alifax.com - info.udine@pec.alifax.com

Entered in Padova companies register at the n. 04337640280. Company stock 10.000.000 entirely deposited. Company with single partner and subjected to direction and coordination of Alifax Holding S.pA.

www.alifax.com