

Certificate of Analysis

PRODUCT DESCRIPTION: NADAL® COVID-19 IgG/IgM plus Test

PRODUCT	LOT	LOT	EXPIRATION	PHYSICAL
CODE	NUMBER	AMOUNT	DATE	APPEARANCE
243004N-25	174501	500000	2022-04	Conforms

PRODUCT TEST RESULTS

TEST	STANDARDS	QC CRITERION	RESULTS	PASS (Y/N)
Background	Whole Blood	Negative	Negative	Yes
Specificity	Negative serum/plasma	Negative	Negative	Yes
Sensitivity	Low Positive IgG	Positive	Positive	Yes
	High Positive IgG	Positive	Positive	Yes
	Low Positive IgM	Positive	Positive	Yes
	High Positive IgM	Positive	Positive	Yes

PRODUCT MANUFACTURING:

No radioactive material of any kind is utilized in the product, or in the manufacturing of the product. The product specified above does not contain any virus, reagent by-product of the same, or metabolic by-product of Hepatitis A, B, C, or D.

If this product is shipped as bulk, nal von minden GmbH assumes no responsibility for the packaging and label claims.

Accepted:

Sandra von Minden

Qualitätsmanagement
nal von minden GmbH



nal von minden GmbH
 Carl-Zeiss-Straße 12
 47445 Moers
 Tel: +49 2841 99820-0
 Fax: +49 2841 99820-1
 info@nal-vonminden.de

Moers, 2021-02-03

nal von minden GmbH

Carl-Zeiss-Straße 12
47445 Moers
Deutschland

Friedenstraße 32
93053 Regensburg
Deutschland

Tel.: +49 941 29010-0
Fax: +49 941 29010-50

info@nal-vonminden.com
www.nal-vonminden.com

Geschäftsführer:
Sandra von Minden
Roland Meißner
Thomas Zander

Registergericht Kleve:
HRB 5679
Steuer-Nr. 244/133/00130
UST-ID-Nr. DE 189 016 086

Sparkasse Regensburg
Kto: 840 017 073
BLZ: 750 500 00

Deutsche Bank
Kto: 056 998 800
BLZ: 350 700 24

Schnelltests

Labordiagnostika

Laborservice

Beratung & Service