

ELI.H.A Aspergillus (102 tests)

Référence / Catalog number **44602**

Lot de conditionnement / Packaging lot : **9152**

Date d'expiration / Expiration date : **2017-12-31**

Réactifs / Reagents	Lot de fabrication / Production lot	Date d'expiration / Expiration date
Flacon R1 (2,2 mL) / R1 (2.2 mL) vial	9152	2017-12-31
Flacon R2 (1 mL) / R2 (1 mL) vial	9152	2017-12-31
Flacon CONTROL + (0,2 mL) / CONTROL + (0.2 mL) vial	9152	2017-12-31
Flacon CONTROL - (0,2 mL) / CONTROL - (0.2 mL) vial	9152	2017-12-31
Flacon R3 (2 mL) / R3 (2 mL) vial	9152	2017-12-31
Flacon BUF (55 mL) / BUF (55 mL) vial	9152	2017-12-31

Certificat de contrôle établi suivant les procédures de contrôles internes
Control certificate established according to the internal control procedures

	Résultats / Results	Résultats théoriques / Theoretical results
R1 Flacon d'Hématies sensibilisées Vial of Sensitized red blood cells		
Conditionnement primaire / Primary packaging	Conforme / Conform to specifications	Conformité / spécifications Conformity / specifications
Etiquetage / Labelling	Conforme / Conform to specifications	Conformité / spécifications Conformity / specifications
Aspect du milieu / Medium appearance	Conforme / Conform to specifications	Suspension d'hématies marron Brown suspension of red blood cells
Performances / Performances	Conforme / Conform to specifications	Sérums humains frais / Fresh human sera Sérums positifs / Positive sera Conformité / spécifications Conformity / specifications
R2 Flacon d'Hématies non sensibilisées Vial of Unsensitized red blood cells		
Conditionnement primaire / Primary packaging	Conforme / Conform to specifications	Conformité / spécifications Conformity / specifications
Etiquetage / Labelling	Conforme / Conform to specifications	Conformité / spécifications Conformity / specifications
Aspect du milieu / Medium appearance	Conforme / Conform to specifications	Suspension d'hématies marron Brown suspension of red blood cells
Performances / Performances	Conforme / Conform to specifications	Sérums humains frais / Fresh human sera Sérums positifs / Positive sera Conformité / spécifications Conformity / specifications
CONTROL +		

	Résultats / Results	Résultats théoriques / Theoretical results
Flacon de Contrôle positif <i>Vial of Positive control</i>		
Conditionnement primaire / <i>Primary packaging</i>	Conforme / Conform to specifications	Conformité / spécifications <i>Conformity / specifications</i>
Etiquetage / <i>Labelling</i>	Conforme / Conform to specifications	Conformité / spécifications <i>Conformity / specifications</i>
Aspect du milieu / <i>Medium appearance</i>	Conforme / Conform to specifications	Liquide jaunâtre translucide <i>Translucent yellowish liquid</i>
Titre / <i>Titer</i>	1 / 320 Conforme / Conform to specifications	Conformité / spécifications <i>Conformity / specifications</i>
CONTROL - Flacon de Contrôle négatif <i>Vial of Negative control</i>		
Conditionnement primaire / <i>Primary packaging</i>	Conforme / Conform to specifications	Conformité / spécifications <i>Conformity / specifications</i>
Etiquetage / <i>Labelling</i>	Conforme / Conform to specifications	Conformité / spécifications <i>Conformity / specifications</i>
Aspect du milieu / <i>Medium appearance</i>	Conforme / Conform to specifications	Liquide jaunâtre translucide <i>Translucent yellowish liquid</i>
R3 Flacon d'adsorbant <i>Vial of Adsorbent</i>		
Conditionnement primaire / <i>Primary packaging</i>	Conforme / Conform to specifications	Conformité / spécifications <i>Conformity / specifications</i>
Etiquetage / <i>Labelling</i>	Conforme / Conform to specifications	Conformité / spécifications <i>Conformity / specifications</i>
Aspect du milieu / <i>Medium appearance</i>	Conforme / Conform to specifications	Suspension d'hématies marron <i>Brown suspension of red blood cells</i>
BUF Flacon de tampon phosphate <i>Vial of Phosphate buffer</i>		
Conditionnement primaire / <i>Primary packaging</i>	Conforme / Conform to specifications	Conformité / spécifications <i>Conformity / specifications</i>
Etiquetage / <i>Labelling</i>	Conforme / Conform to specifications	Conformité / spécifications <i>Conformity / specifications</i>
Aspect du milieu / <i>Medium appearance</i>	Conforme / Conform to specifications	Liquide incolore translucide <i>Translucent colorless liquid</i>
Contrôle du conditionnement / Conformity after packaging	Conforme / Conform to specifications	Conformité / spécifications <i>Conformity / specifications</i>

Référence / *Catalog number* : **44602**
Lot de conditionnement / *Packaging lot* : **9152**

Lot accepté le / *Accepted on* : 13/07/2016

Par / *By* : L.ROBIN
Responsable Contrôle Qualité
Quality Manager