



Certificate of Quality

Product Name: VERSATREK REDOX 1, 40 ML W/STIR BAR
Lot Number: 486796

Product Number: 7106-44
Expiration Date: 2015-10-09
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Pseudomonas aeruginosa	27853	Positive within 96 Hours
Staphylococcus aureus	25923	Positive within 96 Hours
Streptococcus pneumoniae	6305	Positive within 96 Hours

Physical Characteristics:

Appearance: Light Amber, Clear Broth w/ No Significant Precipitate
pH @ 20-25°C: 7.2 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor