



Certificate of Quality

Product Name: VERSATREK REDOX 1, 40 ML W/STIR BAR
Lot Number: 494392

Product Number: 7106-44
Expiration Date: 2015-10-28
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

| Organism Challenge | ATCC | Performance Criteria |
|--------------------------|-------|--------------------------|
| Pseudomonas aeruginosa | 27853 | Positive within 96 Hours |
| Staphylococcus aureus | 25923 | Positive within 96 Hours |
| Streptococcus pneumoniae | 6305 | Positive within 96 Hours |

Physical Characteristics:

Appearance: Light Amber, Clear Broth w/ No Significant Precipitate
pH @ 20-25°C: 7.2 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed

Quality Assurance Supervisor