

## Certificate of Quality

**Product Name:** VERSATREK REDOX 1, 40 ML W/STIR BAR  
**Lot Number:** 625872

**Product Number:** 7106-44  
**Expiration Date:** 2016-08-02  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Quality Control specifications and were found to meet performance criteria for this product.

<b>Organism Challenge</b>	<b>ATCC</b>	<b>Performance Criteria</b>
Pseudomonas aeruginosa	27853	Positive within 96 Hours
Staphylococcus aureus	25923	Positive within 96 Hours
Streptococcus pneumoniae	6305	Positive within 96 Hours

### **Physical Characteristics:**

Appearance: Light Amber, Clear Broth w/ No Significant Precipitate  
pH @ 20-25°C: 7.2 +/- 0.2

### **Microbial Load Evaluation:**

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor