

Certificate of Quality

Product Name: VERSATREK REDOX 2, EZ DRAW 40ML
Lot Number: 983930

Product Number: 7107-44
Expiration Date: 2021-05-27
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Organism Challenge	ATCC	Performance Criteria
Bacteroides fragilis	25285	Positive within 96 Hours
Streptococcus pneumoniae	6305	Positive within 96 Hours

Physical Characteristics:

Appearance: Light Amber, Clear Broth w/ No Pink-Red Discoloration or Significant Precipitate
pH @ 20-25°C: 6.9 +/- 0.2

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Quality Assurance Supervisor