

Certificate of Analysis

Product name: NADAL® RPR Carbon Test

product code	LOT number	expiration date
795010	252	2023-05

Physical Characters

Reagent	Test	Specifications	Results
Latex	Appearance	Liquid suspension, free of macroscopic particles. Grey colour	Pass
Control (+)		Liquid solution. Clear and transparent	
Control (-)		Liquid solution. Clear and transparent	

Functionality Test

Reagent	Test	Specifications	Results
Control (+)	Checking reagent against kit (+) and (-) control	Reactive (R)	R
Control (-)		Negative (N)	N

This lot was tested by test methods described in the manufacturer's package insert.

The individual units comprising lot or lots components from human origin, have been tested for HBsAg and antibodies to HIV (1/2) and HCV and found negative /non-reactive by licensed test in accordance with standard EN 13641. However, handle cautiously as potentially infectious.

Accepted:

Sandra von Minden
Qualitätsmanagement
nal von minden GmbH



nal von minden GmbH
Carl-Zeiss-Straße 12
47445 Moers
Tel: +49 2841 99820-0
Fax: +49 2841 99820-1
info@nal-vonminden.de

Moers, 2021-08-09

nal von minden GmbH

Carl-Zeiss-Str. 12
47445 Moers
Germany

Phone: +49 941 29010-0
Fax: +49 941 29010-50

info@nal-vonminden.com
www.nal-vonminden.com

CEO:
Sandra von Minden
Roland Meißner
Thomas Zander

Bank: Sparkasse Regensburg (GERMANY)
Kto.: 8400 170 73 BLZ.:750 500 00
IBAN: DE9875 0500 0008 4001 7073
BIC/SWIFT: BYLADEM1RBG

Commercial register:
Kleve, HRB 5679
VAT-ID: DE189 016086
Tax-No: 244/133/00130

Rapid Tests

Laboratory Diagnostics

Laboratory Service

Consulting & Service