

QUALITY CONTROL CERTIFICATE

ECONAZOLE 10 µg Neo-Sensitabs

REF: 81312

LOT: 1901-1

Date of analysis: 2019-02

Date of expiry: 2024-02

Appearance: Complies

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

Candida albicans ATCC 90028

Candida parapsilosis ATCC 22019

Candida krusei ATCC 6258

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

ROSCO DIAGNOSTICA A/S



Peter Samuelsen
Quality Assurance Manager