

## QUALITY CONTROL CERTIFICATE

### **KETOCONAZOLE 15 µg Neo-Sensitabs**

REF: 81912N

LOT: 1401-1

Date of analysis: 06-2014

Date of expiry: 06-2019

Appearance: Light yellow, plane tablet, 9 mm in diameter,  
on both sides printed with a black identification  
code.  
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.  
Complies.

#### Performance:

Test strains:

*Candida albicans* ATCC64550  
*Candida albicans* ATCC 90028  
*Candida parapsilosis* ATCC 22019  
*Candida krusei* ATCC 6258

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets  
vary less than  $\pm 1$  mm from mean.  
Complies.

Microbiological quality: Complies.

ROSCO DIAGNOSTICA A/S



José B. Casals  
Quality Assurance Manager

r