

QUALITY CONTROL CERTIFICATE

NYSTATIN 50 µg Neo-Sensitabs

REF: 82212N

LOT: 1301-1

Date of analysis: 06-2013

Date of expiry: 06-2018

Appearance: Light yellow, plane tablet, 9 mm in diameter,
on both sides printed with a black identification
code.
Complies.

Thickness: Mean of 10: 1.50-1.65 mm.
Complies.

Performance:

Test strains:

Candida albicans ATCC 64548


Candida albicans ATCC 10231

Product complies with specifications.

Uniformity: Inhibition zones of individual tablets
vary less than ± 1 mm from mean.
Complies.

Microbiological quality: Complies.

ROSCO DIAGNOSTICA A/S



José B. Casals
Quality Assurance Manager