

## • Certificate of Analysis

Product Name : **Bruker Bacterial Test Standard**

Part No. 8255343  
 LOT No.: 0000279133  
 Use by<sup>1)</sup>: 2018-08  
 Date of Production: 2017-09-14


Storage Conditions : Store below -18 °C immediately on arrival.

<u>Test</u>	<u>Specification</u>	<u>LOT Results</u>
Verification of the calibration points by MALDI-TOF MS	Within a mass range of $\pm 900$ ppm	Passed
Relative intensity of reference proteins No. 3 and No. 4 by MALDI-TOF MS	$0.20 \pm 0.05$	Passed
Side chain reaction	Less or equal than 20% according to MALDI-TOF analysis	Passed
MALDI Biotyper identification	Classification Result Report lists "Escherichia coli" as best match.	Passed
	log(score) Value $\geq$ 2.2 of the best match	Passed

**Remarks:**

<sup>1)</sup> BTS is shipped at ambient temperature but must be stored at -18°C or below immediately on arrival. The expiry date on the package is valid for the enclosed BTS when stored at -18°C or below on arrival. Frozen, dissolved Bruker BTS can be stored for up to 5 months at -18°C or below. See instruction for use at [www.bruker.com/care](http://www.bruker.com/care).

Bruker Daltonik GmbH, Germany  
 Fahrenheitstr. 4 – 28359 Bremen  
 phone +49-421-2205-0

  
**Bruker Daltonik GmbH**  
 Fahrenheitstraße 4  
 D-28359 Bremen  
 Tel: +49 (0) 421 2205-0  
 Dr. M. Kostrzewa, SAFETY OFFICER