

# • Certificate of Analysis

Product Name : **Bruker Matrix HCCA, portioned**

Part No. **8255344**  
 LOT No.: **0000238739**  
 Use by<sup>1)</sup>: **2017-12**  
 Date of Production: **2016-01-14**


Storage Conditions : **Store at 2°C to 8°C on arrival.**

<u>Test</u>	<u>Specification</u>	<u>LOT Results</u>
Differential weighing	Tolerance: ± 0.3 mg	Passed
Functional Check against Protein Calibration Standard I (Bruker)	Signal/Noise: ≥ 100 Resolution: ≥ 250	Passed Passed

**Remarks:**

<sup>1)</sup> Bruker HCCA is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival. Dissolved Bruker HCCA is stable at controlled room temperature (20–25°C) for up to one week.

Bruker Daltonik GmbH, Germany  
 Fahrenheitstr. 4 – 28359 Bremen  
 phone +49-421-2205-0



*B. Schubert*  
 B. Schubert  
 Quality Assurance

QC release date: 15/01/2016

Not for use in diagnostic procedures. For research use only.