

• Certificate of Analysis

Product Name : **Bruker Matrix HCCA, portioned**

Part No. **8255344**
LOT No.: **0000235022**
Use by¹⁾: **2017-10**
Date of Production: **2015-11-20**

Storage Conditions : **Store at 2°C to 8°C on arrival.**

<u>Test</u>	<u>Specification</u>	<u>LOT Results</u>
Differential weighing	Tolerance: ± 0.3 mg	Passed
Functional Check against Protein Calibration Standard I (Bruker)	Signal/Noise: ≥ 100 Resolution: ≥ 250	Passed Passed

Remarks:

¹⁾ Bruker HCCA is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival. Dissolved Bruker HCCA is stable at controlled room temperature (20–25°C) for up to one week.

Bruker Daltonik GmbH, Germany
Fahrenheitstr. 4 – 28359 Bremen
phone +49-421-2205-0



H. Tegtmeyer
i.A. Hannah Tegtmeyer
Quality Assurance

QC release date: 20/11/2015

Not for use in diagnostic procedures. For research use only.