

# • Certificate of Analysis


Product Name: **Bruker Matrix HCCA, portioned**  
 Part No.: 8255344  
 LOT No.: 0000268816  
 Use by<sup>1)</sup>: 2019-04  
 Date of Production: 2017-05-30  
 Storage Conditions : **Store at 2°C to 8°C on arrival.**

<u>Test:</u>	<u>Specification</u>	<u>LOT Results</u>
<b>1. Amount of HCCA content.</b> Sampling in accordance to DIN ISO 3951-1.	AQL-value=1.0 Inspection level II U=2.8 mg // L=2.2 mg	Passed
<b>2. MALDI Biotyper identification of a 1:20 diluted BTS.</b> Sampling in accordance to DIN ISO 2859-1.	Classification Result Report lists "Escherichia coli" as best match.	Passed
<b>3. Blank value analysis of HCCA-portioned by MALDI Biotyper.</b> Sampling in accordance to DIN ISO 2859-1.	AQL-value=1.0 Inspection level S4 L= 2.000 log(score) of the best match	Passed
	Classification Result Report lists "no reliable identification".	Passed
	AQL-value=1.0 Inspection level S4 U= 1.699 log(score)	Passed

**Remarks:**

1) Bruker HCCA is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonik GmbH, Germany  
 Fahrenheitstr. 4 – 28359 Bremen  
 phone +49-421-2205-0

  
**Bruker Daltonik GmbH**  
 Fahrenheitstraße 4  
 D-28359 Bremen  
 Tel: +49 (0) 421 - 22 05 - 0

Dr.M. Kostrzewa, SAFETY OFFICER

QC release date: 07-08-2017

Not for use in diagnostic procedures. For research use only.