

## Certificate of Analysis

Product Name:

**IVD Matrix HCCA-portioned** 

Part No.

8290200

LOT No.:

0000331327

Use by<sup>1)</sup>:

2021-10

**Date of Production:** 

2019-11-07

**Storage Conditions:** 

Store in a refrigerator at 2°C to 8°C on arrival.

Test:	<b>Specification</b>	LOT Results
1. Amount of HCCA content	AQL-value=1.0 Inspection level II U=2.8 mg // L=2.2 mg Sampling in compliance of DIN ISO 3951-1	Passed
2. MALDI Biotyper identification of a 1:20 diluted IVD BTS	Classification Result Report lists "Escherichia coli" as best match with a log score value ≥2.000.	Passed
3. Blank value analysis of IVD matrix HCCA-portioned by MALDI Biotyper	Classification Result Report lists "not reliable identification." with a log score value $\leq 1.699$ .	Passed

**Remarks:** <sup>1)</sup> IVD Matrix HCCA-portioned is shipped at ambient temperature. The expiry date on the package is valid for the product when stored in a refrigerator at 2°C to 8°C on arrival.

Bruker Daltonik GmbH, Germany Fahrenheitstr. 4 – 28359 Bremen phone +49-421-2205-0

QC release date: 2020-01-20

Bruker Daltonik GmbH
Fahrenheitstraße, 4
D-26369 Brømen

Dr.M. Kostrzewa, SAFETY OFFICER

19(0) 421/- 2205-0