

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

| Product                         | Batch     | Production date | Expiration date |
|---------------------------------|-----------|-----------------|-----------------|
| Fusidic Acid FU 0.016-256 µg/mL | 022819021 | 28.02.2019      | 2022.02.27      |

### Ref.

92002 – 920020 – 920021

### Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

| Control strains                                | Medium  | Inoculum                   | Incubation                              | Expected Results<br>MIC range (mg/L) | Results<br>MIC(mg/L) |
|--|---|----------------------------|---|--------------------------------------|----------------------|
| <i>Staphylococcus aureus</i><br>ATCC® 29213    | Mueller Hinton II<br>Agar                     | 0.5 McFarland<br>in saline | 35 ± 2°C, ambient<br>16-20 h            | 0.06–0.25                            | 0.06                 |
| <i>Streptococcus pneumoniae</i><br>ATCC® 49619 | Mueller Hinton II<br>Agar (Sheep<br>blood 5%) | 0.5 McFarland<br>in broth  | 35 ± 2°C, 5% CO <sub>2</sub><br>20-24 h | 4–32                                 | 16                   |

### Batch Release

Approved

Date

09.04.2019

Signature

Quality Control  
(D. Vitagliano)



The results reported were obtained at the time of release.