

Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
Fusidic Acid FU 0.016-256 µg/mL	053121076	31.05.2021	2024.05.30

Ref.

92002 – 920020 – 920021

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.06–0.25	0.125
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO ₂ 20-24 h	4–32	8

Batch Release

Approved

Date

08.06.2021

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.