

Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
Gentamicin CN 0.064-1024 µg/mL	090919019	09.09.2019	2022.09.08

Ref.

92010 – 920100 – 920101

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.12–1	0.19
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	4–16	8
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.25–1	0.38
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.5–2	1.5
<i>Enterococcus faecalis</i> ATCC® 51299	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	> 500	>500

Batch Release

Approved

Date

23.09.2019

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.