

Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
Teicoplanin TEC 0.016-256 µg/mL	122216005	22.12.2016	2019.12.22

Ref.

92012 – 920120 – 920121

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.25–1	0.38
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.25–1	0.75

Batch Release

Approved

Date

02.01.2017

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.