

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Spectinomycin SPC 0.064-1024 µg/mL	012221028	22.01.2021	2024.01.22

Ref.

92014 – 920140 – 920141

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Neisseria gonorrhoeae</i> ATCC® 49226	Mueller Hinton Chocolate Agar	0.5 McFarland in broth	36 ± 1°C, 5% CO ₂ 20-24 h	8–32	8

Batch Release

Approved

Date

02.02.2021

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.