

Liofilchem®

Certificate of Analysis

Page 1 of 1

| Product | Batch | Production date | Expiration date |
|------------------------------|-----------|-----------------|-----------------|
| Oxacillin OX 0.016-256 µg/mL | 013118040 | 31.01.2018 | 2021.01.30 |

Ref.

92015 – 920150 – 920151

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

| Control strains | Medium | Inoculum | Incubation | Expected Results MIC range (mg/L) | Results MIC(mg/L) |
|---|---------------------------|----------------------------|------------------------------|--------------------------------------|----------------------|
| <i>Staphylococcus aureus</i> ATCC® 29213 | Mueller Hinton II Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 16-20 h | 0.12–0.5 | 0.38 |
| <i>Enterococcus faecalis</i> ATCC® 29212 | Mueller Hinton II Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 16-20 h | 8–32 | 12 |

Batch Release

Approved

Date

19.03.2018

Signature

Quality Control
(D. Vitagliano)

The results reported were obtained at the time of release.

Dario Vitagliano