

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Mecillinam MEC 0.016-256 µg/mL	030121036	01.03.2021	2024.02.29

Ref.

92017 – 920170 – 920171

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.03–0.25	0.125

Batch Release

Approved

Date

09.06.2021

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.