

Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
Amoxicillin*-clavulanic acid (2/1) AUG 0.016-256* µg/mL	022217082	22.02.2017	2020.02.22

Ref.	Initial
92024 – 920240 – 920241	AUG

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.12–0.5	0.25
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.25–1	0.5
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	2–8	4
<i>Escherichia coli</i> ATCC® 35218	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	4–16	4
<i>Klebsiella pneumoniae</i> ATCC® 700603	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	4–16	6
<i>Haemophilus influenzae</i> ATCC® 49247	Haemophilus Test Agar	0.5 McFarland in broth	35 ± 2°C, 5% CO ₂ 20-24 h	2–16	3
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO ₂ 20-24 h	0.03–0.12	0.047
<i>Bacteroides fragilis</i> ATCC® 25285	Brucella Blood Agar w Hemin and Vitamin K1	1 McFarland in broth	36 ± 1°C, anaerobically, 24-72 h	0.25–1	0.25
<i>Bacteroides thetaiotaomicron</i> ATCC® 29741	Brucella Blood Agar w Hemin and Vitamin K1	1 McFarland in broth	36 ± 1°C, anaerobically, 24-72 h	0.25–1	0.5

Batch Release

Approved

Date 06.03.2017

Signature

Quality Control
(D. Vitagliano)

The results reported were obtained at the time of release.

Dario Vitagliano