

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Temocillin TMO 0.064-1024 µg/mL	071817034	18.07.2017	2020.07.17

Ref.

92029 – 920290 – 920291

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h		4
<i>Klebsiella pneumoniae</i> ATCC® 700603	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h		12

Batch Release

Approved

Date

02.08.2017

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.