

Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
Kanamycin K 0.016-256 µg/mL	070221019	02.07.2021	2024.07.01

Ref.

92034 – 920340 – 920341

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar*	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4	1
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar*	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	16–64	32
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar*	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4	1.5

*Mueller Hinton II Agar (Cation Adjusted): Liofilchem ref. 10031; BD-BBL ref. 254032.

Batch Release

Approved

Date

14.07.2021

Signature

Quality Control

(D. Vitagliano)



The results reported were obtained at the time of release.