

Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
Clarithromycin CLR 0.016-256 µg/mL	012017099	20.01.2017	2020.01.20

Ref.

92048 – 920480 – 920481

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology


Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.12–0.5	0.19
<i>Haemophilus influenzae</i> ATCC® 49247	Haemophilus Test Agar	0.5 McFarland in broth	35 ± 2°C, ambient 20-24 h	4–16	8
<i>Haemophilus influenzae</i> ATCC® 49247	Haemophilus Test Agar	0.5 McFarland in broth	35 ± 2°C, 5% CO ₂ 20-24 h	8–32	16
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, ambient 20-24 h	0.032–0.125	0.064
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO ₂ 20-24 h	0.064–0.25	0.125

Batch Release

Approved

Date

06.02.2017

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.