

Liofilchem®

Certificate of Analysis

Page 1 of 1

| Product | Batch | Production date | Expiration date |
|------------------------------------|-----------|-----------------|-----------------|
| Clarithromycin CLR 0.016-256 µg/mL | 041718149 | 17.04.2018 | 2021.04.16 |

Ref.

92048 – 920480 – 920481

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology


| Control strains | Medium | Inoculum | Incubation | Expected Results MIC range (mg/L) | Results MIC(mg/L) |
|--|---|----------------------------|---|--------------------------------------|----------------------|
| <i>Staphylococcus aureus</i> ATCC® 29213 | Mueller Hinton II Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 16-20 h | 0.12–0.5 | 0.25 |
| <i>Haemophilus influenzae</i> ATCC® 49247 | Haemophilus Test Agar | 0.5 McFarland in broth | 35 ± 2°C, ambient 20-24 h | 4–16 | 4 |
| <i>Haemophilus influenzae</i> ATCC® 49247 | Haemophilus Test Agar | 0.5 McFarland in broth | 35 ± 2°C, 5% CO ₂ 20-24 h | 8–32 | 8 |
| <i>Streptococcus pneumoniae</i> ATCC® 49619 | Mueller Hinton II Agar (Sheep blood 5%) | 0.5 McFarland in broth | 35 ± 2°C, ambient 20-24 h | 0.032–0.125 | 0.047 |
| <i>Streptococcus pneumoniae</i> ATCC® 49619 | Mueller Hinton II Agar (Sheep blood 5%) | 0.5 McFarland in broth | 35 ± 2°C, 5% CO ₂ 20-24 h | 0.064–0.25 | 0.094 |

Batch Release

Approved

Date

05.06.2018

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.