

Liofilchem®

Certificate of Analysis

Page 1 of 1

| Product | Batch | Production date | Expiration date |
|--------------------------------------|-----------|-----------------|-----------------|
| Vancomycin VA 0.016-256 µg/mL | 020617006 | 06.02.2017 | 2020.02.06 |

Ref.

92057 – 920570 – 920571

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

| Control strains | Medium | Inoculum | Incubation | Expected Results MIC range (mg/L) | Results MIC(mg/L) |
|--|---|----------------------------|---|--------------------------------------|----------------------|
| <i>Staphylococcus aureus</i> ATCC® 29213 | Mueller Hinton II Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 16-20 h | 0.5–2 | 0.75 |
| <i>Enterococcus faecalis</i> ATCC® 29212 | Mueller Hinton II Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 16-20 h | 1–4 | 2 |
| <i>Streptococcus pneumoniae</i> ATCC® 49619 | Mueller Hinton II Agar (Sheep blood 5%) | 0.5 McFarland in broth | 35 ± 2°C, 5% CO ₂ 20-24 h | 0.12–0.5 | 0.25 |

Batch Release

Approved

Date

13.02.2017

Signature

Quality Control
(D. Vitagliano)

Dario Vitagliano

The results reported were obtained at the time of release.