

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Vancomycin VA 0.016-256 µg/mL	091820065	18.09.2020	2023.09.18

Ref.

92057 – 920570 – 920571

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.5–2	0.75
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4	2
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO ₂ 20-24 h	0.12–0.5	0.19

Batch Release

Approved

Date

29.09.2020

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.