

# Liofilchem®

## Certificate of Analysis

Product	Batch	Production date	Expiration date
Vancomycin VA 0.016-256 µg/mL	070721040	07.07.2021	2024.07.06

**Ref.**

92057 – 920570 – 920571

**Antimicrobial Susceptibility Testing**

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.5–2	0.75
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4	3
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO <sub>2</sub> 20-24 h	0.12–0.5	0.5

**Batch Release**

Approved

**Date**

26.07.2021

**Signature**Quality Control  
(D. Vitagliano)

The results reported were obtained at the time of release.