

Liofilchem®

Certificate of Analysis

Page 1 of 1

| Product | Batch | Production date | Expiration date |
|--------------------------------|-----------|-----------------|-----------------|
| Fosfomicin FOS 0.016-256 µg/mL | 051817023 | 18.05.2017 | 2020.05.17 |

Ref.

92078 – 920780 – 920781

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

| Control strains | Medium | Inoculum | Incubation | Expected Results MIC range (mg/L) | Results MIC(mg/L) |
|--|---------------------------|----------------------------|------------------------------|--------------------------------------|----------------------|
| <i>Staphylococcus aureus</i> ATCC® 29213 | Mueller Hinton II Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 16-20 h | 0.5–4 | 0.5 |
| <i>Enterococcus faecalis</i> ATCC® 29212 | Mueller Hinton II Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 16-20 h | 32–128 | 32 |
| <i>Escherichia coli</i> ATCC® 25922 | Mueller Hinton II Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 16-20 h | 0.5–2 | 0.75 |
| <i>Pseudomonas</i> <i>aeruginosa</i> ATCC® 27853 | Mueller Hinton II Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 16-20 h | 2–8 | 3 |

Batch Release

Approved

Date

24.05.2017

Signature

Quality Control
(D. Vitagliano)

Dario Vitagliano

The results reported were obtained at the time of release.