

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Metronidazole MTZ 0.016-256 µg/mL	042618051	26.04.2018	2021.04.25

Ref.

92087 – 920870 – 920871

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Bacteroides fragilis</i> ATCC® 25285	Brucella Blood Agar w Hemin and Vitamin K1	1 McFarland in broth	36 ± 1°C, anaerobically, 24-72 h	0.25–2	0.5
<i>Bacteroides thetaiotaomicron</i> ATCC® 29741	Brucella Blood Agar w Hemin and Vitamin K1	1 McFarland in broth	36 ± 1°C, anaerobically, 24-72 h	0.5–4	1
<i>Eggerthelia lenta</i> ATCC® 43055	Brucella Blood Agar w Hemin and Vitamin K1	1 McFarland in broth	36 ± 1°C, anaerobically, 24-72 h	0.125–0.5	0.125

Batch Release

Approved

Date

15.06.2018

Signature

Quality Control
(D. Vitagliano)

Dario Vitagliano

The results reported were obtained at the time of release.