

Liofilchem®

Certificate of Analysis

Page 1 of 1

| Product | Batch | Production date | Expiration date |
|-----------------------------------|-----------|-----------------|-----------------|
| Metronidazole MTZ 0.016-256 µg/mL | 012020020 | 20.01.2020 | 2023.01.19 |

Ref.

92087 – 920870 – 920871

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

| Control strains | Medium | Inoculum | Incubation | Expected Results MIC range (mg/L) | Results MIC(mg/L) |
|--|--|-------------------------|--|--------------------------------------|----------------------|
| <i>Bacteroides fragilis</i> ATCC® 25285 | Brucella Blood Agar w Hemin and Vitamin K1 | 1 McFarland in broth | 36 ± 1°C, anaerobically, 24-72 h | 0.25–2 | 0.38 |
| <i>Bacteroides thetaiotaomicron</i> ATCC® 29741 | Brucella Blood Agar w Hemin and Vitamin K1 | 1 McFarland in broth | 36 ± 1°C, anaerobically, 24-72 h | 0.5–4 | 2 |
| <i>Eggerthelia lenta</i> ATCC® 43055 | Brucella Blood Agar w Hemin and Vitamin K1 | 1 McFarland in broth | 36 ± 1°C, anaerobically, 24-72 h | 0.125–0.5 | 0.19 |

Batch Release

Approved

Date

26.02.2020

Signature

Quality Control
(D. Vitagliano)

Dario Vitagliano

The results reported were obtained at the time of release.