

Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
Netilmicin NET 0.016-256 µg/mL	072717072	27.07.2017	2020.07.26

Ref.

92093 – 920930 – 920931

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	≤0.25	0.25
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	4-16	8
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	≤0.5-1	0.5
<i>Pseudomonas</i> <i>aeruginosa</i> ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.5-8	2

Batch Release

Approved

Date

04.09.2017

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.