

# Liofilchem®

## Certificate of Analysis

Product	Batch	Production date	Expiration date
<b>Penicillin G P 0.016-256 µg/mL</b>	080219039	02.08.2019	2022.08.01

**Ref.**

**92102 – 921020 – 921021**

**Antimicrobial Susceptibility Testing**

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.25–2	0.25
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4	3
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO <sub>2</sub> 20-24 h	0.25–1	0.38
<i>Neisseria gonorrhoeae</i> ATCC® 49226	Mueller Hinton Chocolate Agar	0.5 McFarland in broth	36 ± 1°C, 5% CO <sub>2</sub> 20-24 h	0.25–1	0.25
<i>Bacteroides fragilis</i> ATCC® 25285	Brucella Blood Agar w Hemin and Vitamin K1	1 McFarland in broth	36 ± 1°C, anaerobically, 24-72 h	8–32	12
<i>Bacteroides thetaiotaomicron</i> ATCC® 29741	Brucella Blood Agar w Hemin and Vitamin K1	1 McFarland in broth	36 ± 1°C, anaerobically, 24-72 h	8–32	12

**Batch Release**

Approved

**Date**

09.09.2019

**Signature**

Quality Control  
(D. Vitagliano)

*Dario Vitagliano*

The results reported were obtained at the time of release.