

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Penicillin G P 0.016-256 µg/mL	100720034	07.10.2020	2023.10.07

Ref.

92102 – 921020 – 921021

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.25–2	0.25
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4	3
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO ₂ 20-24 h	0.25–1	0.38
<i>Neisseria gonorrhoeae</i> ATCC® 49226	Mueller Hinton Chocolate Agar	0.5 McFarland in broth	36 ± 1°C, 5% CO ₂ 20-24 h	0.25–1	0.5
<i>Bacteroides fragilis</i> ATCC® 25285	Brucella Blood Agar w Hemin and Vitamin K1	1 McFarland in broth	36 ± 1°C, anaerobically, 24-72 h	8–32	32
<i>Bacteroides thetaiotaomicron</i> ATCC® 29741	Brucella Blood Agar w Hemin and Vitamin K1	1 McFarland in broth	36 ± 1°C, anaerobically, 24-72 h	8–32	16

Batch Release

Approved

Date

19.10.2020

Signature

Quality Control
(D. Vitagliano)

Dario Vitagliano

The results reported were obtained at the time of release.