

# Liofilchem®

## Certificate of Analysis

Product	Batch	Production date	Expiration date
<b>Piperacillin*-tazobactam TZP 0.016-256* µg/mL</b>	120919025	09.12.2019	2022.12.08

**Ref.**

92108 – 921080 – 921081

**Antimicrobial Susceptibility Testing**

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.25–2	1
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4	3
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–4	2
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	1–8	3
<i>Escherichia coli</i> ATCC® 35218	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.5–2	1
<i>Klebsiella pneumoniae</i> ATCC® 700603	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	8–32	12
<i>Haemophilus influenzae</i> ATCC® 49247	Haemophilus Test Agar	0.5 McFarland in broth	35 ± 2°C, 5% CO <sub>2</sub> 20-24 h	0.06–0.5	0.25
<i>Bacteroides fragilis</i> ATCC® 25285	Brucella Blood Agar w Hemin and Vitamin K1	1 McFarland in broth	36 ± 1°C, anaerobically, 24-72 h	0.12–0.5	0.38
<i>Bacteroides thetaiotaomicron</i> ATCC® 29741	Brucella Blood Agar w Hemin and Vitamin K1	1 McFarland in broth	36 ± 1°C, anaerobically, 24-72 h	2–16	16
<i>Eggerthelia lenta</i> ATCC® 43055	Brucella Blood Agar w Hemin and Vitamin K1	1 McFarland in broth	36 ± 1°C, anaerobically, 24-72 h	8–32	8

**Batch Release**

Approved

**Date**

02.01.2020

**Signature**

Quality Control  
(D. Vitagliano)

*Dario Vitagliano*

The results reported were obtained at the time of release.