

# Liofilchem®

## Certificate of Analysis

Product	Batch	Production date	Expiration date
Streptomycin S 0.064-1024 µg/mL	040220030	02.04.2020	2023.04.02

**Ref.**

92111 – 921110 – 921111

**Antimicrobial Susceptibility Testing**

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar*	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	<500	128
<i>Enterococcus faecalis</i> ATCC® 51299	Mueller Hinton II Agar*	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	≥1000	>1024

\*Mueller Hinton II Agar (Cation Adjusted): Liofilchem ref. 10031; BD-BBL ref. 254032.

**Batch Release**

Approved

**Date**

09.04.2020

**Signature**Quality Control  
(D. Vitagliano)

The results reported were obtained at the time of release.