

Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
Streptomycin S 0.064-1024 µg/mL	020521010	05.02.2021	2024.02.05

Ref.

92111 – 921110 – 921111

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar*	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	<500	96
<i>Enterococcus faecalis</i> ATCC® 51299	Mueller Hinton II Agar*	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	≥1000	>1024

*Mueller Hinton II Agar (Cation Adjusted): Liofilchem ref. 10031; BD-BBL ref. 254032.

Batch Release

Approved

Date

17.02.2021

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.