

# Liofilchem®

## Certificate of Analysis

Product	Batch	Production date	Expiration date
<b>Trimethoprim*-sulfamethoxazole (1/19) SXT 0.002-32* µg/mL</b>	042320050	23.04.2020	2023.04.23

**Ref.**

92123 – 921230 – 921231

**Antimicrobial Susceptibility Testing**

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	≤0.5	0.064
<i>Enterococcus faecalis</i> ATCC® 29212	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	≤0.5	0.032
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	≤0.5	0.125
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	8–32	32
<i>Haemophilus influenzae</i> ATCC® 49247	Haemophilus Test Agar	0.5 McFarland in broth	35 ± 2°C, 5% CO <sub>2</sub> 20-24 h	0.03–0.25	0.19
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO <sub>2</sub> 20-24 h	0.12–1	0.38
<i>Haemophilus influenzae</i> ATCC® 49766	Mueller Hinton Fastidious Agar	0.5 McFarland in saline	35 ± 1°C, 5% CO <sub>2</sub> 16-20 h	0.016–0.06*	0.06

\*Established and validated by EUCAST

**Batch Release**

Approved

Date 11.05.2020

Signature

Quality Control  
(D. Vitagliano)

The results reported were obtained at the time of release.