

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
Cefuroxime CXM 0.016-256 µg/mL	041420058	14.04.2020	2023.04.14

Ref.

92129 – 921290 – 921291

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.5–2	1
<i>Escherichia coli</i> ATCC® 25922	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	2–8	6
<i>Haemophilus influenzae</i> ATCC® 49766	Haemophilus Test Agar	0.5 McFarland in broth	35 ± 2°C, 5% CO ₂ 20-24 h	0.25–1	0.5
<i>Streptococcus pneumoniae</i> ATCC® 49619	Mueller Hinton II Agar (Sheep blood 5%)	0.5 McFarland in broth	35 ± 2°C, 5% CO ₂ 20-24 h	0.25–1	0.25
<i>Neisseria gonorrhoeae</i> ATCC® 49226	Mueller Hinton Chocolate Agar	0.5 McFarland in broth	36 ± 1°C, 5% CO ₂ 20-24 h	0.25–1	0.38

Batch Release

Approved

Date

29.04.2020

SignatureQuality Control
(D. Vitagliano)

The results reported were obtained at the time of release.