

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
FLUCONAZOLE FLU 0.016-256 µg/mL	021417087	2020.02.14
Ref.	Initial	
92147 – 921470 – 921471	FLU	

Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Candida krusei</i> ATCC® 6528	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	128–>256	256
<i>Candida parapsilosis</i> ATCC® 22019	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	1–8	2
<i>Candida albicans</i> ATCC® 90028	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.125–0.5	0.38

Batch Release

Approved

Date

22.02.2017

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.