

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
<b>Fluconazole FLU</b> <b>0.002-32 µg/mL</b>	022318078	23.02.2018	2021.02.22

**Ref.**

92147 – 921470 – 921471

**Antimicrobial Susceptibility Testing**

Tested according to in-house methodology for antifungal agents


Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Candida krusei</i> ATCC® 6258	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	128–≥256	128
<i>Candida parapsilosis</i> ATCC® 22019	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	1–8	3
<i>Candida albicans</i> ATCC® 90028	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.125–0.5	0.5

**Batch Release**

Approved

**Date**

16.04.2018

**Signature**Quality Control  
(D. Vitagliano)

The results reported were obtained at the time of release.