

# Liofilchem®

## Certificate of Analysis

Page 1 of 1

Product	Batch	Production date	Expiration date
FLUCONAZOLE FLU 0.016-256 µg/mL	022119026	21.02.2019	2022.02.20

### Ref.

92147 – 921470 – 921471

### Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Candida krusei</i> ATCC® 6258	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	128–≥256	>256
<i>Candida parapsilosis</i> ATCC® 22019	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	1–8	3
<i>Candida albicans</i> ATCC® 90028	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.125–0.5	0.125

### Batch Release

Approved

Date

05.03.2019

Signature

Quality Control  
(D. Vitagliano)



The results reported were obtained at the time of release.