

Liofilchem®

Certificate of Analysis

Page 1 of 1

| Product | Batch | Production date | Expiration date |
|---------------------------------|-----------|-----------------|-----------------|
| FLUCONAZOLE FLU 0.016-256 µg/mL | 040819051 | 08.04.2019 | 2022.04.07 |

Ref.

92147 – 921470 – 921471

Antimicrobial Susceptibility Testing

Tested according to in-house methodology for antifungal agents

| Control strains | Medium | Inoculum | Incubation | Expected Results MIC range (mg/L) | Results MIC(mg/L) |
|--|---------------|----------------------------|------------------------------|--------------------------------------|----------------------|
| <i>Candida krusei</i> ATCC® 6258 | R.P.M.I. Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 24-48 h | 128–≥256 | >256 |
| <i>Candida parapsilosis</i> ATCC® 22019 | R.P.M.I. Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 24-48 h | 1–8 | 3 |
| <i>Candida albicans</i> ATCC® 90028 | R.P.M.I. Agar | 0.5 McFarland in saline | 35 ± 2°C, ambient 24-48 h | 0.125–0.5 | 0.38 |

Batch Release

Approved

Date

29.04.2019

Signature

Quality Control
(D. Vitagliano)



The results reported were obtained at the time of release.