

# Liofilchem®

## Certificate of Analysis

Product	Batch	Production date	Expiration date
FLUCONAZOLE FLU 0.016-256 µg/mL	022621002	26.02.2021	2024.02.26

**Ref.**

92147 – 921470 – 921471

**Antimicrobial Susceptibility Testing**

Tested according to in-house methodology for antifungal agents

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Candida krusei</i> ATCC® 6258	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	128–>256	128
<i>Candida parapsilosis</i> ATCC® 22019	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	1–8	4
<i>Candida albicans</i> ATCC® 90028	R.P.M.I. Agar	0.5 McFarland in saline	35 ± 2°C, ambient 24-48 h	0.125–0.5	0.19

**Batch Release**

Approved

**Date**

15.03.2021

**Signature**Quality Control  
(D. Vitagliano)

The results reported were obtained at the time of release.